



Analytical Method Information

Analyte	CAS #	DL	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Blank Spike / LCS %R	RPD
Turbidity, SM 2130 B (SM 2130 B-11) in Water								
Preservation: Cool <6°C								
Container: Small OJ, 500 mL								
Minimum Sample Volume: 250 mL								
Hold Time: 2 days								
Turbidity		0.0500 NTU	0.0500 NTU	0.0500 NTU	20		90 - 110	20
Turbidity, SM 2130 B (SM 2130 B-11) in Drinking Water								
Preservation: Cool <6°C								
Container: Small OJ, 500 mL								
Minimum Sample Volume: 250 mL								
Hold Time: 2 days								
Turbidity		0.0500 NTU	0.0500 NTU	0.0500 NTU	20		90 - 110	20