



**Analytical Method Information**

Analyte	CAS #	DL	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Matrix Spike RPD	Blank Spike / LCS %R	Blank Spike / LCS RPD
<b>Conductivity, Specific Conductance, 120.1 (EPA 120.1) in Water</b>									
Preservation: Cool <6°C									
Container: Small OJ, 500 mL									
Minimum Sample Volume: 500 mL									
Hold Time: 28 days									
Conductivity		1.00 µS/cm	1.00 µS/cm	1.00 µS/cm	20			90 - 110	
<b>Conductivity, Specific Conductance, 120.1 (EPA 120.1) in Drinking Water</b>									
Preservation: Cool <6°C									
Container: Small OJ, 500 mL									
Minimum Sample Volume: 500 mL									
Hold Time: 28 days									
Conductivity		1.00 µS/cm	1.00 µS/cm	1.00 µS/cm					