



Analytical Method Information

Analyte	CAS #	DL	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Matrix Spike RPD	Blank Spike / LCS %R	Blank Spike / LCS RPD
Iron, Ferrous SM 3500-Fe B (SM 3500-Fe B-11) in Water									
Preservation: HCL									
Container: Glass NM, Amber, 500 mL, HCL									
Minimum Sample Volume: 500 mL									
Hold Time: 1 day									
Ferrous Iron		0.0400 mg/L	0.0400 mg/L	0.0400 mg/L	20	75 - 125		90 - 110	20