



Analytical Method Information

Analyte	CAS #	DL	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Matrix Spike RPD	Blank Spike / LCS %R	Blank Spike / LCS RPD
Chloride, EPA 325.2 (EPA 325.2) in Water									
Preservation: None									
Container: Small OJ, 500 mL									
Minimum Sample Volume: 500 mL									
Hold Time: 28 days									
Chloride	16887-00-6	1.00 mg/L	1.00 mg/L	1.00 mg/L	20	75 - 125	20	90 - 110	20