



### Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Blank Spike / LCS %R	RPD
<b>Chromium, Hexavalent, SM 3500-Cr B-09 (SM 3500-Cr B-09) in Water</b>								
Preservation:	Cool <6°C							
Container:	Glass NM, Amber, 250 mL		Minimum Sample Volume: 250 mL		Hold Time: 1 day			
Hexavalent Chromium	0.0100 mg/L	0.0100 mg/L	0.0100 mg/L		20	85 - 115	85 - 115	20