



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R RPD	Blank Spike / LCS %R RPD
Organic Carbon, Total, 9060A (EPA 9060A) in Water							
Preservation: pH<2; H2SO4, Cool <6°C							
Container: Glass NM, Amber, 250 mL, 9N H2SO4		Minimum Sample Volume: 250 mL		Hold Time: 28 days			
Total Organic Carbon	0.500 mg/L	0.500 mg/L	0.500 mg/L		20	75 - 125	90 - 110 20