



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R RPD	Blank Spike / LCS %R RPD
pH, SM 4500-H (SM 4500-H+ B-00) in Water							
Preservation: None							
Container: Small OJ, 500 mL		Minimum Sample Volume: 250 mL		Hold Time: 0.01 days			
pH	0.0100 pH Units	0.0100 pH Units	0.0100 pH Units		20		