



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Blank Spike / LCS %R	RPD
8270D 1,4-Dioxane (Low Level H₂O or Solid) (EPA 8270D) in Water								
Preservation: Cool <6°C								
Container:	Glass NM, Amber, 500 mL			Minimum Sample Volume: 1000 ml	Hold Time: 7 days			
1,4-Dioxane	0.179 ug/L	0.200 ug/L	0.400 ug/L		30	35.1 - 120	30	39.9 - 120
surr: 1,4-Dioxane-d8				33.6 - 120				
1,4-Dichlorobenzene-d4								