



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Blank Spike / LCS %R	RPD
8270D-SIM Butyl Tins (EPA 8270D-SIM) in Water								
Preservation: Cool <6°C								
Container:	Glass NM, Amber, 500 mL		Minimum Sample Volume: 200 mL	Hold Time: 7 days				
Tributyltin Ion	0.0430 ug/L	0.0960 ug/L	0.193 ug/L	30	30 - 160	30	30 - 160	30
Dibutyltin Ion	0.0960 ug/L	0.145 ug/L	0.289 ug/L	30	30 - 160	30	30 - 160	30
Butyltin Ion	0.108 ug/L	0.204 ug/L	0.204 ug/L	30	30 - 160	30	30 - 160	30
Tetrabutyltin	0.300 ug/L	0.300 ug/L	0.300 ug/L					30
surr: Tripentyltin			30 - 160					
surr: Tripropyltin			30 - 160					
Tetrapentyltin								
p-Terphenyl-d14								