



### Analytical Method Information

---

**Analysis: Turbidity, EPA 180.1 (EPA 180.1) in Water**

**Preservation:** Cool <6°C

**Container:** Small OJ, 500 mL

**Minimum Sample Volume:** 250 mL

**Hold Time:** 2 days

---

Analyte	CAS #	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Matrix Spike RPD	Blank Spike / LCS %R	Blank Spike / LCS RPD
Turbidity		0.0500 NTU	0.0500 NTU	0.0500 NTU		20			90 - 110	20

---



### Analytical Method Information

---

**Analysis: Turbidity, SM 2130 B-01 (SM 2130 B-01) in Water**

**Preservation:** Cool <6°C

**Container:** Small OJ, 500 mL

**Minimum Sample Volume:** 250 mL

**Hold Time:** 2 days

---

Analyte	CAS #	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Matrix Spike RPD	Blank Spike / LCS %R	Blank Spike / LCS RPD
Turbidity		0.0500 NTU	0.0500 NTU	0.0500 NTU		20			90 - 110	20

---