



Analytical Method Information

Analyte	CAS #	DL	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Matrix Spike RPD	Blank Spike / LCS %R	Blank Spike / LCS RPD
pH, EPA 9040C (EPA 9040C) in Water									
Preservation: None									
Container: Small OJ, 500 mL									
Minimum Sample Volume: 250 mL									
Hold Time: 0.01 days									
pH		0.0100 pH Units	0.0100 pH Units	0.0100 pH Units	20			99.2 - 100.8	