



### Analytical Method Information

Analyte	CAS #	DL	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Matrix Spike RPD	Blank Spike / LCS %R	Blank Spike / LCS RPD
<b>Met 200.8 - As (EPA 200.8) in Water</b>									
Preservation: pH<2; HNO <sub>3</sub> , Cool <6°C									
Container: HDPE NM, 500 mL, 1:1 HNO <sub>3</sub>									
Minimum Sample Volume: 500 mL									
Hold Time: 180 days									
Arsenic-75a	7440-38-2	0.0326 ug/L	0.100 ug/L	0.200 ug/L	20	75 - 125	20	80 - 120	20
Arsenic-75b	7440-38-2	0.0602 ug/L	0.250 ug/L	0.500 ug/L	20	75 - 125	20	80 - 120	20