



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R RPD	Blank Spike / LCS %R RPD
Conductivity, Specific Conductance, 120.1 (EPA 120.1) in Water							
Preservation: Cool <6°C							
Container: Small OJ, 500 mL							
Minimum Sample Volume: 500 mL							
Hold Time: 28 days							
Conductivity	1.00 µS/cm	1.00 µS/cm	1.00 µS/cm		20		90 - 110