



Analytical Method Information

| Analyte | DL | LOD | LOQ / RL | Surrogate %R | Duplicate RPD | Matrix Spike %R | RPD | Blank Spike / LCS %R | RPD |
|---|-------------|------------|------------|-----------------|------------------|--------------------|-----|-------------------------|-----|
| 8270D-SIM Butyl Tins (EPA 8270D-SIM) in Tissue | | | | | | | | | |
| Preservation: Freeze <10°C | | | | | | | | | |
| Container: Glass WM, Clear, 8 oz | | | | | | | | | |
| Minimum Sample Weight: 200 mL | | | | | | | | | |
| Hold Time: 365 days | | | | | | | | | |
| Tributyltin Ion | 0.450 ug/kg | 1.93 ug/kg | 3.86 ug/kg | | 35 | 20 - 130 | 35 | 20 - 130 | 35 |
| Dibutyltin Ion | 1.73 ug/kg | 4.33 ug/kg | 5.78 ug/kg | | 35 | 20 - 130 | 35 | 20 - 130 | 35 |
| Butyltin Ion | 1.89 ug/kg | 3.06 ug/kg | 4.08 ug/kg | | 35 | 20 - 130 | 35 | 20 - 130 | 35 |
| Tetrabutyltin | 5.00 ug/kg | 5.00 ug/kg | 5.00 ug/kg | | | | | | 35 |
| surr: Tripentyltin | | | | | 20 - 130 | | | | |
| surr: Tripropyltin | | | | | 20 - 130 | | | | |
| Tetrapentyltin | | | | | | | | | |
| p-Terphenyl-d14 | | | | | | | | | |