



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R RPD	Blank Spike / LCS %R RPD
Chloride, EPA 325.2 (EPA 325.2) in Water							
Preservation: None							
Container: Small OJ, 500 mL		Minimum Sample Volume: 500 mL		Hold Time: 28 days			
Chloride	1.00 mg/L	1.00 mg/L	1.00 mg/L		20	75 - 125	90 - 110 20