



### Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
<b>Chromium, Hexavalent, 7196A (EPA 7196A) in Water</b>									
Preservation: Cool <6°C									
Container: Glass NM, Amber, 250 mL									
Minimum Sample Volume: 250 mL									
Hold Time: 1 day									
Hexavalent Chromium	0.0100 mg/L	0.0100 mg/L	0.0100 mg/L		20	85 - 115		85 - 115	20