



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R RPD	Blank Spike / LCS %R RPD
Iron, Ferrous SM 3500-Fe B-97 (SM 3500-Fe B-97) in Water							
Preservation: HCL							
Container: Glass NM, Amber, 500 mL, HCL		Minimum Sample Volume: 500 mL		Hold Time: 1 day			
Ferrous Iron	0.0400 mg/L	0.0400 mg/L	0.0400 mg/L		20	75 - 125	90 - 110 20