



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
Met 7470A Hg Low Level (EPA 7470A) in Water									
Preservation: pH<2; HNO ₃ , Cool <6°C									
Container: HDPE NM, 500 mL, 1:1 HNO ₃									
Minimum Sample Volume: 500 mL									
Hold Time: 28 days									
Mercury	0.00000260 mg/L	0.0000100 mg/L	0.0000200 mg/L		20	75 - 125	20	80 - 120	20