



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
RSK-175 Dissolved Gases (MEE) (EPA RSK-175) in Water									
Preservation: None									
Container: VOA Vial, Clear, 40 mL									
Minimum Sample Volume: 80 mL									
Hold Time: 14 days									
Methane	0.340 ug/L	0.654 ug/L	0.654 ug/L		30	80 - 120	30	80 - 120	30
Ethane	0.390 ug/L	0.615 ug/L	1.23 ug/L		30	80 - 120	30	80 - 120	30
Ethene	0.240 ug/L	0.570 ug/L	1.14 ug/L		30	80 - 120	30	80 - 120	30
Acetylene	0.550 ug/L	1.06 ug/L	1.06 ug/L		30	73 - 123	30	73 - 123	30
surr: Propane					72 - 122	62 - 122	30	62 - 122	30