



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R RPD	Blank Spike / LCS %R RPD
Turbidity, EPA 180.1 (EPA 180.1) in Water							
Preservation: Cool <6°C							
Container: Small OJ, 500 mL							
Minimum Sample Volume: 250 mL							
Hold Time: 2 days							
Turbidity	0.050 NTU	0.050 NTU	0.050 NTU		20		90 - 110 20