



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
608 PCB 0.01 ug/L or 20 ug/kg (EPA 608) in Water									
Preservation: Cool <6°C									
Container: Glass NM, Amber, 1000 mL									
Minimum Sample Volume: 2000 mL									
Hold Time: 7 days									
Aroclor 1016	0.00248 ug/L	0.00500 ug/L	0.0100 ug/L		30	44 - 117	30	44 - 117	30
Aroclor 1221	0.00276 ug/L	0.00500 ug/L	0.0100 ug/L		30				
Aroclor 1232	0.00276 ug/L	0.00500 ug/L	0.0100 ug/L		30				
Aroclor 1242	0.00248 ug/L	0.00500 ug/L	0.0100 ug/L		30				
Aroclor 1248	0.00248 ug/L	0.00500 ug/L	0.0100 ug/L		30				
Aroclor 1254	0.00248 ug/L	0.00500 ug/L	0.0100 ug/L		30				
Aroclor 1260	0.00276 ug/L	0.00500 ug/L	0.0100 ug/L		30	46 - 131	30	46 - 131	30
Aroclor 1262	0.00276 ug/L	0.00500 ug/L	0.0100 ug/L		30				
Aroclor 1268	0.00276 ug/L	0.00500 ug/L	0.0100 ug/L		30				
surr: Decachlorobiphenyl				21 - 100					
surr: Tetrachlorometaxylene				19 - 111					
1-Bromo-2-Nitrobenzene									
Hexabromobiphenyl									