



**Analytical Method Information**

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
<b>608 PCB 1.0 ug/L (EPA 608) in Water</b>									
Preservation: Cool <6°C									
Container: Glass NM, Amber, 500 mL									
Minimum Sample Volume: 1000 ml									
Hold Time: 7 days									
Aroclor 1016	0.130 ug/L	0.500 ug/L	1.00 ug/L		30	62 - 120	30	62 - 120	30
Aroclor 1221	0.147 ug/L	0.500 ug/L	1.00 ug/L		30				
Aroclor 1232	0.147 ug/L	0.500 ug/L	1.00 ug/L		30				
Aroclor 1242	0.147 ug/L	0.500 ug/L	1.00 ug/L		30				
Aroclor 1248	0.130 ug/L	0.500 ug/L	1.00 ug/L		30				
Aroclor 1254	0.130 ug/L	0.500 ug/L	1.00 ug/L		30				
Aroclor 1260	0.147 ug/L	0.500 ug/L	1.00 ug/L		30	61 - 120	30	61 - 120	30
Aroclor 1262	0.147 ug/L	0.500 ug/L	1.00 ug/L		30				
Aroclor 1268	0.147 ug/L	0.500 ug/L	1.00 ug/L		30				
surr: Decachlorobiphenyl					29 - 120				
surr: Tetrachlorometaxylene					35 - 120				
1-Bromo-2-Nitrobenzene									
Hexabromobiphenyl									