



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	Blank Spike / LCS %R	RPD
8041A Chlorinated Phenols (EPA 8041A) in Water								
Preservation: Cool <6°C								
Container:	Glass NM, Amber, 500 mL		Minimum Sample Volume: 1000 mL	Hold Time: 7 days				
2,4-Dichlorophenol	1.51 ug/L	3.00 ug/L	6.00 ug/L	30	30 - 160	30	30 - 160	30
2,4,6-Trichlorophenol	0.106 ug/L	0.130 ug/L	0.250 ug/L	30	30 - 160	30	30 - 160	30
2,3,6-Trichlorophenol	0.0660 ug/L	0.130 ug/L	0.250 ug/L	30	30 - 160	30	30 - 160	30
2,4,5-Trichlorophenol	0.0770 ug/L	0.130 ug/L	0.250 ug/L	30	30 - 160	30	30 - 160	30
2,3,4-Trichlorophenol	0.126 ug/L	0.130 ug/L	0.250 ug/L	30	30 - 160	30	30 - 160	30
2,3,4,6- and 2,3,5,6-Tetrachlorophenol	0.0910 ug/L	0.130 ug/L	0.250 ug/L	30	30 - 160	30	30 - 160	30
2,3,4,5-Tetrachlorophenol	0.0770 ug/L	0.130 ug/L	0.250 ug/L	30	30 - 160	30	30 - 160	30
Pentachlorophenol	0.0850 ug/L	0.130 ug/L	0.250 ug/L	30	48 - 120	30	48 - 120	30
surr: 2,4,6-Tribromophenol				26 - 120				