



Analytical Method Information

Analyte	DL	LOD	LOQ / RL	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
Met 200.8 - Cu (EPA 200.8) in Water									
Preservation: pH<2; HNO ₃ , Cool <6°C									
Container: HDPE NM, 500 mL, 1:1 HNO ₃									
Minimum Sample Volume: 500 mL									
Hold Time: 180 days									
Copper-63	0.378 ug/L	0.378 ug/L	0.500 ug/L		20	75 - 125	20	80 - 120	20
Copper-65	0.350 ug/L	0.350 ug/L	0.500 ug/L		20	75 - 125	20	80 - 120	20