



Analytical Method Information

| Analyte | DL | LOD | LOQ / RL | Surrogate %R | Duplicate RPD | Matrix Spike %R RPD | Blank Spike / LCS %R RPD |
|--|------------|------------|------------|-----------------|------------------|------------------------|-----------------------------|
| Turbidity, EPA 180.1 (EPA 180.1) in Water | | | | | | | |
| Preservation: Cool <6°C | | | | | | | |
| Container: Small OJ, 500 mL | | | | | | | |
| Minimum Sample Volume: 250 mL | | | | | | | |
| Hold Time: 2 days | | | | | | | |
| Turbidity | 0.0500 NTU | 0.0500 NTU | 0.0500 NTU | | 20 | | 90 - 110 20 |



Analytical Method Information

| Analyte | DL | LOD | LOQ / RL | Surrogate %R | Duplicate RPD | Matrix Spike %R RPD | Blank Spike / LCS %R RPD |
|--|------------|------------|------------|-----------------|------------------|------------------------|-----------------------------|
| Turbidity, SM 2130 B-01 (SM 2130 B-01) in Water | | | | | | | |
| Preservation: Cool <6°C | | | | | | | |
| Container: Small OJ, 500 mL | | | | | | | |
| Minimum Sample Volume: 250 mL | | | | | | | |
| Hold Time: 2 days | | | | | | | |
| Turbidity | 0.0500 NTU | 0.0500 NTU | 0.0500 NTU | | 20 | | 90 - 110 20 |