

SUMMARY OF QC CRITERIA FOR USE WITH 6020D

Quality Control Type	Minimum Frequency	Acceptance Criteria	Difference
Initial Calibration Blank	Daily	Must not contain target analytes above half the LLOQ	Currently 1X the LLOQ
Internal Standards	Continuous	70-125%	Currently 60-125%

LLOQ-Lower Limit of Quantitation. In most cases this is the lowest concentration in the calibration curve.